

1

2

3

4

REV	ECO	Release	Drawn	REVISION HISTORY
-0	2022-033	2022-11-07	SGO	Customer Drawing - Insert



Model 310-104
p/n 00REG-1034-02

GENERAL SPECIFICATIONS

Inlet Pressure Rating:	3000 PSI MAX
Regulated Output:	16 ± 2 PSI (Dynamic) (NOM 1 Bar) 20 ± 2 PSI (Static)
Average Open Flow Rate:	50 L/min (AVG)
Weight:	3.2 Oz. [92 g]
Spare O-ring:	MH p/n 09001-3113-70 (DIN-477-9N)
See also:	MH document 5SREG-310-xxx

Installing the Regulator

Seat the inlet nipple of the regulator into the corresponding outlet socket of the cylinder valve and turn the grip nut to engage the valve threads.

Complete the connection by turning the grip nut **HAND TIGHT ONLY! DO NOT use a wrench or pliers. Over-tightening will damage the regulator.** The integrity of the connection is provided by an o-ring seal and is not dependant on the tightness of the threaded coupling.

Connect the outlet tubing and assemble the remainder of your oxygen system (EDS unit or Flowmeter).

Open the cylinder valve **SLOWLY** (~ 2 turns).

Removing the Regulator


DO NOT ATTEMPT TO REMOVE REGULATOR WHILE UNDER PRESSURE

The regulator grip-nut will be difficult to turn while under pressure, and doing so will destroy the regulator inlet O-ring.

Bleed-off internal pressure by first **closing the main cylinder valve** and then disconnecting the outlet tubing from the regulator.

The grip nut should now turn easily **by hand** and the regulator can be removed.

OUTLET TUBING CONNECTION - 4mm/6mm Push-To-Connect Fittings



INSERTING

Push in the tubing


REMOVING

1. Push in the connector collar
2. Pull the tube straight back while holding the collar in

To INSERT TUBING: Push the tubing into the connector until resistance is felt, then push a little further, about **1/8 inch [3 mm]**. Gently tug on the tubing to make sure it is captured.

To REMOVE TUBING: Push the tubing in slightly, then push in the connector collar while pulling gently on the tubing.

When removing tubing, **DO NOT pull on the tubing without pushing in the collar**, as this will likely damage the connector.




Note: Alternate right-angle outlet fitting provided.

- Replace installed fitting as preferred.
- Maintain cleanliness for oxygen service.
- No Teflon tape required
- Torque to 20 in-lbs.

UNLESS OTHERWISE SPECIFIED,
DIMENSIONS ARE IN INCHES.
DIMENSIONS IN [] ARE MILLIMETERS (REF)
TOLERANCES ARE:
0.X ±0.015 ANGLES ±3°
0.XX ±0.010 FRACTIONS ±1/64
0.XXX ±0.005
63
INTERPRET GD&T PER ASME 14.5

THIRD ANGLE
PROJECTION



DO NOT SCALE
DRAWING

DRAWN SGO
2022-09-27

CHECKED EAM
2022-11-04

ENGINEER JB
2022-11-04

APPROVED HBS
2022-11-07

MH MOUNTAIN HIGH E&S CO. REDMOND, OR. USA			
THIS DOCUMENT AND ALL TECHNICAL DATA HEREON DISCLOSED ARE PROPERTY OF MOUNTAIN HIGH E&S CO. AND SHALL NOT BE USED, RELEASED OR DISCLOSED IN WHOLE OR PART WITHOUT WRITTEN PERMISSION FROM MOUNTAIN HIGH E&S CO. THIS DOCUMENT MUST BE RETURNED TO MOUNTAIN HIGH E&S CO. IMMEDIATELY UPON REQUEST.			
DWG TITLE MH 3G Regulator, Single-Stage, NO-Gauge, DIN-477-N x 4mm Axial Tube (XCR) [Insert]		DWG REV. -0	
DWG NUMBER 5IREG-1034-02		DWG REV. -0	
SRC FILE 00REG-310-xxx\$-0	INV. PART NUMBER 00REG-1034-02	PROD. NAME	
DWG FORMAT: ESR-002 Rev H [27]	DWG SCALE	DWG SHEET 1 OF 1	DWG SIZE A 11x8½

Insert #: 5IREG-1034-02

1

2

3